SAFE HANDLING GUIDE

# Carbon-14 Handling Precautions

This document contains general information designed to provide a basic understanding of radiation safety. While we believe the information to be accurate, regulatory requirements may change and information contained herein is not tailored to individual needs. A radiation protection specialist should be consulted for specific applications.

<sup>14</sup>C 5730 y β- 0.156 No γ

E 0.156

## **Physical data**

Maximum beta energy: 0.156 MeV (100%)<sup>(1)</sup>
Maximum range of beta in air: 22 cm (8.6 in)<sup>(2)</sup>

### Occupational limits(3)

Annual limit on intake: 2 mCi (74 MBq)

Derived air concentration: 1 x 10<sup>-6</sup> µCi/mL (37 kBq/m<sup>3</sup>)

#### **Dosimetry**

Millicurie (37 MBq) quantities of <sup>14</sup>C do not present a significant external exposure hazard because the low-energy betas emitted barely penetrate the outer dead layer of skin. <sup>14</sup>C-labeled compound uptake may be assumed to be uniformly distributed throughout all organs and tissues in the body<sup>(4)</sup>. Most <sup>14</sup>C-labeled compounds are rapidly metabolized and the radionuclide is exhaled as <sup>14</sup>CO<sub>2</sub>. Some compounds and their metabolites are eliminated via the urine. Biological half lives vary from a few minutes to 40 days<sup>(4)</sup>.

## **General handling precautions for Carbon-14**

- Designate area for handling <sup>14</sup>C and clearly label all containers.
- 2. Prohibit eating, drinking, smoking and mouth pipetting in room where <sup>14</sup>C is handled.
- 3. Use transfer pipets, spill trays and absorbent coverings to confine contamination.
- 4. Handle potentially volatile compounds in ventilated enclosures.
- 5. If enhanced containment is necessary, handle volatile compounds in closed systems vented through suitable traps.
- Sample exhausted effluent and room air by drawing a known volume through a membrane filter followed by an impinger containing dilute NaOH.
- 7. Wear disposable lab coats, wrist guards and gloves for secondary protection.
- 8. Select gloves appropriate for chemicals handled.
- Maintain contamination and exposure control by regularly monitoring and promptly decontaminating gloves and surfaces.
- 10. Use pancake or end-window Geiger-Mueller detectors or liquid scintillation counter to detect <sup>14</sup>C.
- 11. Submit periodic urine and breath samples (as appropriate) for bioassay to determine uptake by personnel.



Some  $^{14}$ C-labeled compounds may penetrate gloves and skin. Handle these compounds remotely, wear two pairs of gloves and change the outer layer frequently. Special caution should be observed when handling  $^{14}$ C-labeled halogenated acids. These compounds can be incorporated in the skin and deliver local dose commitments in the order of 10-100 rad per  $\mu$ Ci (3-30 Gy per MBq) deposited.

#### References

- Kocher, David C., Radioactive Decay Data Tables, Springfield: National Technical Information Service, 1981 DOE/TIC-11026.
- 2. Kaplan, Irving, Nuclear Physics, New York: Addison-Wesley, 1964.
- 3. U.S. Nuclear Regulatory Commission. 10 CFR 20 Appendix B Standards for Protection Against Radiation, 1994.
- 4. ICRP Publication 30, Part 3, Limits for Intakes of Radionuclides by Workers. Pergamon Press, Oxford, 1981.

PerkinElmer, Inc. 940 Winter Street Waltham, MA 02451 USA P: (800) 762-4000 or (+1) 203-925-4602 www.perkinelmer.com

